



**NIEUWKOOP**


# **OPERATING MANUAL**

## **LPR10M0T**

Pyranometer



EN  
V1.0



## Contents

<b>1</b>	<b>Introduction .....</b>	<b>3</b>
<b>2</b>	<b>Technical specifications.....</b>	<b>4</b>
<b>3</b>	<b>Measuring principle .....</b>	<b>6</b>
<b>4</b>	<b>Installation.....</b>	<b>7</b>
4.1	Optional mounting supports.....	8
4.2	Electrical connections .....	9
<b>5</b>	<b>Configuration.....</b>	<b>10</b>
<b>6</b>	<b>ASCII Proprietary protocol .....</b>	<b>11</b>
<b>7</b>	<b>Modbus-RTU protocol .....</b>	<b>15</b>
<b>8</b>	<b>Use of the diagnostic sensors .....</b>	<b>18</b>
<b>9</b>	<b>Maintenance .....</b>	<b>19</b>
<b>10</b>	<b>Safety instructions .....</b>	<b>20</b>
<b>11</b>	<b>Accessories ordering codes .....</b>	<b>21</b>

## 1 Introduction

The pyranometer measures the **global irradiance** on a flat surface ( $\text{W/m}^2$ ), sum of direct solar irradiance and diffuse irradiance.

**LPR10M0T** falls within the **Fast Response Spectrally Flat Class A** pyranometers according to the ISO 9060:2018 standard and meets the requirements of the WMO "Guide to Instruments and Methods of Observation".

The internal temperature, relative humidity and pressure **diagnostic sensors** allow keeping the pyranometer operating conditions under control all the time and foresee any maintenance intervention in advance, thus ensuring always reliable measurements.

The integrated heater prevents dew and frost, ensuring high accuracy in all weather conditions. The heater can operate in automatic mode, turning on and off according to the temperature, or it can be turned on/off on user command.

The integrated bubble level and the adjustable feet facilitate horizontal positioning during installation. The **tilt sensor** allows continuous monitoring of the correct installation.

RS485 Modbus-RTU isolated output.


The pyranometer is factory calibrated in accordance with the ISO 9847:2023 (Type A1) standard: "Calibration of pyranometers by comparison to a reference pyranometer". The calibration is performed by comparison with the reference sample calibrated annually at WRC (World Radiation Center). The pyranometer is supplied with calibration report, including measurement of temperature response and directional response, as required by the standard for Class A pyranometers.

The **DATAsense** PC application software, downloadable from the website, allows configuring the sensor, monitoring the measurements in real time, even graphically for the irradiance, and saving the values detected during the connection in a file.

## 2 Technical specifications

Sensor	Thermopile
Measuring range	-200...4000 W/m <sup>2</sup>
Resolution	0.1 W/m <sup>2</sup>
Viewing angle	2 $\pi$ sr
Spectral range (50%)	285...2850 nm
Output	Isolated RS485 Modbus-RTU
Power supply	7...30 Vdc
Consumption	2 W (heater ON) / 0.3 W (heater OFF)
Connection	5-pole M12
Weight	650 g approx.
Operating conditions	-40...+80 °C / 0...100 %RH / Max. altitude 6000 m
Bubble level accuracy	< 0.2°
Protection degree	IP 67
Materials	Housing: anodized aluminium Screen: ASA Dome: optical glass
MTBF	> 10 years

### Dimensions (mm)



**Technical Specifications According to ISO 9060:2018**

Classification	Fast Response Spectrally Flat Class A
Response time (95%)	< 0.3 s
Zero offset	
a) response to a 200 W/m <sup>2</sup> thermal radiation	<  ±1  W/m <sup>2</sup>
b) response to a 5 K/h change in ambient temperature	<  ±1  W/m <sup>2</sup>
c) total zero offset including the effects a), b) and other sources	<  ±4  W/m <sup>2</sup>
Long-term instability (1 year)	<  ±0.5  %
Non-linearity	<  ±0.2  %
Directional response (up to 80° with 1000 W/m <sup>2</sup> beam)	<  ±10  W/m <sup>2</sup>
Spectral error	<  ±0.2  %
Temperature response (-10...+40°C)	<  ±0.5  %
Tilt response	<  ±0.4  %

**Diagnostic sensors**

<b>Internal Temperature</b>	
Measuring range	-40...+80 °C
Resolution	0.1 °C
Accuracy	±0.5 °C (0...60 °C)
<b>Internal Relative Humidity</b>	
Measuring range	0...100%
Resolution	0.1%
Accuracy	±3% @ T=25 °C & RH=20...80%
<b>Internal pressure</b>	
Measuring range	500...1100 hPa
Resolution	0.1 hPa
Accuracy	±1 hPa (0...60 °C)

**Tilt sensor (only LPS10MxT)**

Measuring range	0...180°
Resolution	0.1°
Accuracy	< 0.5°

### 3 Measuring principle


The pyranometer is based on a thermopile sensor, which allows the pyranometer not to be selective at different wavelengths.

The pyranometer spectral response is determined by the optical glass dome and diffuser. The diffuser optimizes the angular response of the pyranometer.

The glass dome also protects the sensitive area from weathering.

The special material used to manufacture the dome and diffuser allows the spectral range to be extended to short wavelengths starting from 285 nm.

Considering a standard solar spectrum, the portion of solar irradiation detected by the pyranometer is greater than 99.8%. Figure 3.1 shows the relative spectral sensitivity of the LPR10M0T pyranometer (**blue line**) and the standard solar spectrum (**red line**).



**Fig. 3.1: LPR10M0T relative spectral sensitivity and standard solar spectrum**

By sensing the temperature of the sensor, the pyranometer automatically compensates, with a microprocessor circuit, the response of the sensor itself with respect to temperature changes, optimizing the response and minimizing temperature drift.


To prevent condensation from forming on the internal side of the dome under certain climatic conditions, silica-gel is inserted inside the pyranometer to absorb moisture. The efficiency status of the silica-gel can be monitored by detecting the measurement of the relative humidity diagnostic sensor (see the chapter "Use of the diagnostic sensors").

## 4 Installation

- The pyranometer must be mounted in an easy-to-reach location in order to clean the dome regularly and carry out maintenance. At the same time, make sure that no buildings, constructions, trees or obstructions exceed the horizontal plane where the pyranometer lies. If this is not possible, select a site where obstructions in the path of the sun from sunrise to sunset do not exceed 5 degrees of elevation. **N.B.: The presence of obstructions on the horizon line affects significantly the measurement of direct irradiance.**
- The pyranometer must be located far from any kind of obstruction, which might reflect sunlight (or sun shadow) onto the pyranometer itself.
- The mast height does not exceed the pyranometer plane to avoid measurement errors caused by any reflection or shadow of the mast itself.
- In compliance with ISO TR9901 standard and WMO recommendations, if the pyranometer is used without the solar radiation protection screen, it must be positioned so that the connector points to the North Pole, if the instrument is used in the Northern Hemisphere, and to the South Pole, if used in the Southern Hemisphere. In any case, it is better to follow this recommendation even when the screen is used.
- For fixing, use the M5 holes in the lower part of the pyranometer. For an accurate horizontal positioning, adjust the height of the two lower feet with hexagonal ring nut, using the bubble level integrated in the pyranometer.


*Note: for horizontal positioning it is preferable to use the bubble level instead of the tilt sensor, as it is more accurate; use the tilt sensor for positioning monitoring after installation. Use the tilt sensor for installation only if the pyranometer must be positioned tilted with respect to the horizontal plane.*

- It is preferable to thermally insulate the pyranometer from its mounting bracket by securing it not with the base directly in contact with the support plate, but by leaving a layer of air in between (for this purpose, always use the feet even if the installation is not horizontal), while at the same time ensuring that there is good electrical contact to ground.




**Fig. 4.1: position of fixing holes and bubble level**

## 4.1 Optional mounting supports



**Fig. 4.2: LPS1 bracket for mast**



**Fig. 4.3: LPS5 adjustable bracket for mast**





## 4.2 Electrical connections

### Warning!

The metallic housing of the pyranometer should preferably be grounded locally. The CPM12... cable shield is connected to the connector shell and then to the metallic housing of the pyranometer. Do not connect the yellow/green wire (cable shield) to ground, unless it is not possible to ground the pyranometer metallic housing locally via the support mast.

Internally there are surge protection devices connected to the housing. Grounding the housing allows the correct protection functionality of the devices.

Pyranometer male connector (external view)		Function	CPM12-5... wire color
	1	+Vdc (Power supply positive)	Brown
	2	GND (Power supply negative)	White
	3	DATA + (RS485)	Blue
	4	DATA - (RS485)	Black
	5	DGND (Digital ground)	Grey
	Shell	Pyranometer housing / Cable shield	Yellow/Green



**Fig. 4.4: connection diagram**

The RS485 output is isolated. Before connecting the pyranometer to the RS485 network, set the address and the communication parameters, if different from the factory preset (see "Configuration" chapter).

## 5 Configuration

It is possible to set:

- The user code (mnemonic name to identify the pyranometer).
- The serial communication parameters (Modbus address, Baud Rate, parity/stop bit).
- The heating operating mode (Auto, ON, OFF).
- The alarm thresholds of operating time and parameters measured by the diagnostic sensors.
- The temperature unit of measurement (diagnostic sensor).


The configuration of the pyranometer can be done:

- With the **DATAsense** PC application software, downloadable from the website.  
*Note: the application software requires entering the password **9876** to change the configuration of the pyranometer (see the software online help for details).*
- By sending serial commands from a PC, via a standard communication program (see the chapter "ASCII proprietary protocol").
- Via the "Holding Registers" in MODBUS-RTU mode (only serial communication parameters, alarm thresholds and temperature unit of measurement, see the chapter "Modbus-RTU protocol").

By default, the instrument has MODBUS address **1** and communication parameters 19200, 8E1.

### Connection to PC:

To connect the pyranometer to a PC USB port, to check or change the configuration, the optional **CP24B-5** kit can be used.



**Fig. 4.5: CP24K-5 kit**

To use the CP24K-5 kit, the related USB drivers must be installed in the PC.

Alternatively, it is possible to use a standard RS485/USB or RS485/RS232 converter, powering the pyranometer separately (the minimum supply voltage is 7 V).

## 6 ASCII Proprietary protocol

To communicate with the pyranometer via the ASCII proprietary protocol:

1. Connect the pyranometer to the PC and start a standard serial communication program.
2. In the serial communication program, set the Baud Rate 57600, the parameters 8N2 and the COM port number to which the pyranometer is connected.
3. Power the pyranometer (or power cycle if already powered) and send the command @ within 5 seconds from the instrument power on (the pyranometer replies &| if the command @ is recognized).  
**Note:** if the CP24K-5 kit is used, to power cycle the pyranometer, disconnect the power supply jack connector for a few seconds, then reconnect it.
4. Send the command **CAL USER ON** (the pyranometer replies CAL USER ON;USER ON) to enable the configuration change. The command CAL USER ON is not required only for reading the settings.
5. Send the commands described in the tables below.

The command CAL USER ON is automatically disabled after a few minutes of inactivity. The command CAL USER ON can be disabled immediately by sending the command CAL EXIT.

### Ping:

Command	Reply	Description
P0	P0;&	Ping.

### Pyranometer general information:

Command	Reply	Description
G0	G0;model	Model
G1	G1;sub-model	Sub-model
G2	G2;nnnnnnnn	Serial number
G3	G3;xx.yy	Firmware revision
G4	G4;yyyy/mm/dd	Firmware revision date
G5	G5;xx.yy	Hardware revision
G6	G6;nnnnn	Operating time in days from first power on
G7	G7;nnnnn	Operating time in days from last power on
GD	GD;yyyy/mm/dd hh:mm:ss	Factory calibration date and time

### User code:

The user code is an alphanumeric string of up to 8 characters that the user can set to identify the pyranometer with a mnemonic name.

Command	Reply	Description
CUCnnnnnnnn	CUC;&	Sets nnnnnnnn as user code. Default= <i>blank</i>
RUC	RUC;nnnnnnnn	Reads the user code.

**RS485 Modbus-RTU communication parameters:**

Command	Reply	Description
CMA <sub>n</sub>	CMA; <sub>&amp;</sub>	Sets the Modbus-RTU address (1...247) to n. Default=1
RMA	RMA; <sub>n</sub>	Reads the Modbus-RTU address.
CMB <sub>n</sub>	CMB; <sub>&amp;</sub>	Sets the Baud Rate: <ul style="list-style-type: none"> <li>▪ 9600 if n=0</li> <li>▪ 19200 if n=1 (<i>default</i>)</li> <li>▪ 38400 if n=2</li> <li>▪ 57600 if n=3</li> <li>▪ 115200 if n=4</li> </ul>
RMB	RMB; <sub>n</sub>	Reads Baud Rate setting.
CMP <sub>n</sub>	CMP; <sub>&amp;</sub>	Sets parity and stop bits (data bits = 8 fixed): <ul style="list-style-type: none"> <li>▪ 8N1 if n=0</li> <li>▪ 8E1 if n=2 (<i>default</i>)</li> <li>▪ 8O1 if n=4</li> <li>▪ 8N2 se n=1</li> <li>▪ 8E2 se n=3</li> <li>▪ 8O2 se n=5</li> </ul>
RMP	RMP; <sub>n</sub>	Reads the setting of parity and stop bits.
CMW <sub>n</sub>	CMW; <sub>&amp;</sub>	Sets waiting time after transmission with Modbus-RTU protocol: <ul style="list-style-type: none"> <li>▪ Immediate reception if n=0 (violates protocol)</li> <li>▪ Waiting 3.5 characters if n=1 (respects protocol)</li> </ul> <i>Default</i> : Waiting 3.5 characters (n=1)
RMW	RMW; <sub>n</sub>	Reads the setting of waiting time after transmission with Modbus-RTU protocol.

**Heating:**

Heating can be set in automatic mode or always on or always off. In automatic mode, heating is automatically turned on/off according to the detected temperature..

Command	Reply	Description
WHEATER <sub>n</sub>	WHEATER; <sub>&amp;</sub>	Sets the heating operating mode: <ul style="list-style-type: none"> <li>▪ Always OFF if n=0 (<i>default</i>)</li> <li>▪ Always ON if n=1</li> <li>▪ Automatic if n=2</li> </ul>
RHEATER	RHEATER; <sub>n</sub>	Reads the heating operating mode.

**Units of measurement:**

Changing the temperature units of measurement affects the Modbus protocol only. The temperature measurement provided by the ASCII proprietary protocol (commands S0/S1) is always in °C.

Command	Reply	Description
TT <sub>n</sub>	WUT; <sub>&amp;</sub>	Sets the temperature unit of measurement: <ul style="list-style-type: none"> <li>▪ °C if n=0 (<i>default</i>)</li> <li>▪ °F if n=1</li> <li>▪ K if n=2</li> </ul>
RUT	RUT; <sub>n</sub>	Reads the temperature unit of measurement.

**Measurements reading:**

Command	Reply	Description
S0	S0;meas	<p>Prints the measurements in the following sequence:</p> <ul style="list-style-type: none"> <li>▪ Temperature compensated irradiance in W/m<sup>2</sup> (it takes into account the change in sensor sensitivity as temperature changes)</li> <li>▪ Nominal irradiance in W/m<sup>2</sup> (calculated considering the sensor nominal sensitivity at the calibration temperature: ~23 °C)</li> <li>▪ Signal in mV generated by the thermopile</li> <li>▪ Internal temperature (*) in °C</li> <li>▪ Internal relative humidity in %</li> <li>▪ Internal pressure in hPa</li> <li>▪ Tilt angle</li> <li>▪ Power supply voltage in V</li> <li>▪ Operating time in days from first power on</li> <li>▪ Operating time in days from last power on</li> <li>▪ Operating time alarm from first power on (0/1)</li> <li>▪ Operating time alarm from last power on (0/1)</li> <li>▪ Internal temperature alarm (0/1)</li> <li>▪ Internal relative humidity alarm (0/1)</li> <li>▪ Internal pressure alarm (0/1)</li> <li>▪ Reserved</li> <li>▪ Reserved</li> <li>▪ Reserved</li> <li>▪ Sensor temperature (*) in °C</li> <li>▪ Reserved</li> <li>▪ Reserved</li> <li>▪ Current heating status (0=OFF; 1=ON)</li> <li>▪ Heating operating mode (0=Always OFF; 1=Always ON; 2=AUTO)</li> </ul> <p>If automatic sending of the measurements is enabled (command S1), sending S0 disables the automatic sending.</p>
S1	S1;meas	Enables the automatic sending of the measurements, once per second.

(\*) The pyranometer has 2 temperature sensors, one inside the housing, with diagnostic function, and one at sensor level, for heating control in AUTO mode. The two temperatures are generally different.

**Diagnostics alarm thresholds:**

Command	Reply	Description
WAL1n...n	WAL1;&	Sets to n...n the operating time alarm threshold from first power on (0...65535 days, default=730 days → 2 years).
AL1	AL1;n...n	Reads the operating time alarm threshold from first power on.
WAL2n...n	WAL2;&	Sets to n...n the operating time alarm threshold from last power on (0...65535 days, default=730 days → 2 years).
AL2	AL2;n...n	Reads the operating time alarm threshold from last power on.
WAL3n...n	WAL3;&	Sets to n...n, in the set unit of measurement, the internal temperature alarm threshold (0...150 °C, default=80 °C).

Command	Reply	Description
AL3	AL3;n...n	Reads the internal temperature alarm threshold.
WAL4n...n	WAL4;&	Sets to n...n% the internal relative humidity alarm threshold (0...100%, default=25%).
AL4	AL4;n...n	Reads the internal relative humidity alarm threshold.
WAL5n...n	WAL5;&	Sets to n...n hPa the internal pressure alarm threshold (800...1100 hPa, default=1100 hPa).
AL5	AL5;n...n	Reads the internal pressure alarm threshold.

All alarm thresholds are integer values with resolution 1 (1 day, 1 °C/°F/K, 1 %RH, 1 hPa).

The alarm condition occurs when the parameter becomes greater than the set threshold.

## 7 Modbus-RTU protocol

By default, the pyranometer has Modbus address **1** and communication parameters 19200, 8E1. The address and the communication parameters can be changed by using the appropriate serial commands of the proprietary protocol or, alternatively, directly with Modbus commands by changing the value of the Holding Registers described later.

The Modbus-RTU protocol is active after a few seconds from the sensor power on.

The delay between the reception of a query from the master device and the start of the pyranometer reply is less than 10 ms.

Below is the list of registers.

### Input Registers:

Address	Description	Format
<b>MEASUREMENTS</b>		
1 + 2	Temperature compensated irradiance in W/m <sup>2</sup> (x10) <i>(it takes into account the change in sensor sensitivity as temperature changes)</i>	32-bit Integer
3 + 4	Nominal irradiance in W/m <sup>2</sup> (x10) <i>(calculated considering the sensor nominal sensitivity at the calibration temperature: ~23 °C)</i>	32-bit Integer
6	Internal relative humidity in % (x10)	16-bit Integer
7	Internal temperature (*) in the set unit of measurement (x10)	16-bit Integer
8	Internal pressure in hPa (x10)	16-bit Integer
9 + 10	Signal in mV generated by the thermopile (x1000)	32-bit Integer
11	Tilt angle (x10)	16-bit Integer
12	Sensor temperature (*) in the set unit of measurement (x10)	16-bit Integer
13	Current heating status: 0=OFF; 1=ON	16-bit Integer
<b>PYRANOMETER GENERAL INFORMATION</b>		
16...25	Model (20 characters)	String
26...35	Sub-model (20 characters)	String
36...39	Serial number (8 characters)	String
40...43	Firmware revision (8 characters)	String
44...47	Hardware revision (8 characters)	String
<b>CALIBRATION HISTORY</b>		
50 + 51	Sensitivity of last calibration in $\mu\text{V}/\text{Wm}^{-2}$ (x1000)	32-bit Integer
52...55	Date of last calibration (yyyymmdd)	String
56 + 57	Sensitivity of historical calibration 1 in $\mu\text{V}/\text{Wm}^{-2}$ (x1000)	32-bit Integer
58...61	Date of historical calibration 1 (yyyymmdd)	String
62 + 63	Sensitivity of historical calibration 2 in $\mu\text{V}/\text{Wm}^{-2}$ (x1000)	32-bit Integer
64...67	Date of historical calibration 2 (yyyymmdd)	String
68 + 69	Sensitivity of historical calibration 3 in $\mu\text{V}/\text{Wm}^{-2}$ (x1000)	32-bit Integer

Address	Description	Format
70...73	Date of historical calibration 3 (yyyymmdd)	String
74 + 75	Sensitivity of historical calibration 4 in $\mu\text{V}/\text{Wm}^{-2}$ (x1000)	32-bit Integer
76...79	Date of historical calibration 4 (yyyymmdd)	String
80 + 81	Sensitivity of historical calibration 5 in $\mu\text{V}/\text{Wm}^{-2}$ (x1000)	32-bit Integer
82...85	Date of historical calibration 5 (yyyymmdd)	String
<b>OPERATING TIME</b>		
100	Operating time in days from first power on	16-bit Integer
101	Operating time in days from last power on	16-bit Integer

(\*) The pyranometer has 2 temperature sensors, one inside the housing, with diagnostic function, and one at sensor level, for heating control in AUTO mode. The two temperatures are generally different.

### Notes on the registers format:

- 1) To read a **32-bit integer** value, the two consecutive registers indicated must be accessed. The register with lower address contains the most significant bits.

Example: the temperature compensated irradiance, for ex.  $50.1 \text{ W}/\text{m}^2$ , is available in the registers with address 1 and 2.

Register address	1	2
Register content (Hex)	0000	01F5
Read value	$0x000001F5 = 501 = 50.1 \text{ W}/\text{m}^2$	

- 2) The alphanumeric **strings** are contained in a series of 16-bit registers. Each register contains 2 characters of the string: the most significant 8 bits contain the ASCII code of the first character, the least significant 8 bits contain the ASCII code of the next character. The start character of the string is in the register of the series with the lowest address.

Example: the pyranometer model, for ex. LPR10M0T, is available in the registers with address from 16 to 25.

Register address	16	17	18	19	20	21	22	23	24	25
Register content (Hex)	4C50	5231	304D	3054	0000	0000	0000	0000	0000	0000
Read value	L	P	R	1	0	M	A	T		

### Discrete inputs:

Address	Description	Format
0	Operating time alarm from first power on	Bit
1	Operating time alarm from last power on	Bit
2	Internal temperature alarm	Bit
3	Internal relative humidity alarm	Bit
4	Internal pressure alarm	Bit

The value of a "Discrete input", normally at 0, becomes 1 when the alarm threshold set with the relative "Holding register" is exceeded.



**Holding Registers:**

Address	Description	Format
0	RS485 Modbus-RTU Baud Rate: 0=9600 1=19200 (default) 2=38400 3=57600 4=115200	16-bit Integer
1	RS485 Modbus-RTU parity and stop bits: 0=8N1 1=8N2 2=8E1 (default) 3=8E2 4=8O1 5=8O2	16-bit Integer
2	Modbus-RTU address (1...247, default=1).	16-bit Integer
5	Temperature unit of measurement: 0=°C (default) 1=°F 2=K	16-bit Integer
6	Operating time alarm threshold from first power on (0...65535 days, default=730 days → 2 years)	16-bit Integer
7	Operating time alarm threshold from last power on (0...65535 days, default=730 days → 2 years)	16-bit Integer
8	Internal temperature alarm threshold in the set unit of measurement (0...150 °C, default=80 °C)	16-bit Integer
9	Internal relative humidity alarm threshold in % (0...100 %, default=25%)	16-bit Integer
10	Internal pressure alarm threshold in hPa (800...1100 hPa, default=1100 hPa)	16-bit Integer
11	Heating operating mode: 0=Always OFF ( <i>default</i> ) 1=Always ON 2=Automatic (ON/OFF depending on temperature)	16-bit Integer
30002	Enabling writing in Holding Registers: 9876=Enable 8888=Disable	16-bit Integer
30003	Holding Register writing enabling status: 0=Disabled 1=Enabled	16-bit Integer

All alarm thresholds are integer values with resolution 1 (1 day, 1 °C/°F/K, 1 %RH, 1 hPa). To change the settings via the Holding Registers, it is first necessary to enable the writing by entering the value 9876 in the register with address 30002.

## 8 Use of the diagnostic sensors

The internal temperature, relative humidity, pressure and tilt **diagnostic sensors** allow keeping the pyranometer operating conditions under control all the time and foresee any maintenance intervention in advance, thus ensuring always reliable measurements.

### Internal temperature

---

As a rule, the internal temperature of the pyranometer is on average 5 to 10 °C higher than the external ambient temperature. Temperatures that are excessively lower or higher than indicated may be a sign of malfunctions.

Monitoring the internal temperature therefore makes it possible to determine whether the irradiance measurement can be considered reliable.

### Internal relative humidity

---

To minimize condensation and keep measurements accurate, desiccant silica-gel is provided inside the pyranometer base to absorb moisture. The silica-gel life is at least 10 years. The long-term monitoring of the internal relative humidity of the pyranometer allows the efficiency of the silica-gel to be checked. A progressive upward trend in relative humidity indicates the progressively decreasing ability of silica-gel to absorb moisture.

The internal relative humidity depends not only on the saturation level of the silica-gel but also on the temperature of the pyranometer; therefore, short-term humidity monitoring is not particularly significant in determining the condition of the silica-gel. Relative humidity monitoring can be annual.

It is advisable to detect the relative humidity during the night, when the temperature is lower and the relative humidity is higher.

Silica-gel can be considered close to saturation when the internal relative humidity is persistently above 50% at temperatures below 20 °C.

### Internal pressure

---

Monitoring the internal pressure of the pyranometer allows the housing to be checked for tight seal (no leakage). The pressure/temperature ratio should remain approximately constant.

### Tilt

---

Monitoring the inclination angle of the pyranometer installation allows for the detection of unwanted displacements that may occur in the long term or as a result of impacts suffered by the installation.

Take the measurement of the Tilt sensor immediately after installation and consider it as a reference for subsequent measurements.

### Alarm thresholds

---

Thresholds can be defined for internal temperature, relative humidity and pressure. When the thresholds are exceeded an alarm condition is generated, which can be read via the Modbus "Discrete inputs" or the DATAsense software.

The thresholds can be set via the Modbus "Holding registers", the WAL3/WAL4/WAL5 serial commands or the DATAsense software.

## 9 Maintenance

In order to grant measurements high accuracy, it is important to keep the outer glass dome clean. Consequently, the more the dome will be kept clean, the more measurements will be accurate.

You can wash it using water and standard papers for lens. If necessary, use pure ETHYL alcohol. After using alcohol, clean again the dome with water only.

To minimize condensation and keep measurements accurate, desiccant silica-gel is provided inside the pyranometer base to absorb moisture. The silica-gel life is at least 10 years and can be replaced by the user.

To replace the silica-gel, unscrew the 4 screws that fix the base to the main body of the pyranometer; remove the silica-gel tablet and insert a new tablet, then screw the pyranometer base back on.



### **Warning!**

The silica-gel replacement must be done in an ESD-safe environment.

To exploit all the pyranometer features, it is highly recommended that the calibration be checked annually.

The pyranometer provides the operating time in days since first and last power on, and allows time alarm thresholds to be set to alert the user of the need for periodic verification or maintenance. The exceeding of thresholds can be read via Modbus "Discrete inputs" or DATAsense software. The thresholds can be set via Modbus "Holding registers", WAL1/WAL2 serial commands, or DATAsense software.

## 10 Safety instructions

The pyranometer proper operation and operating safety can be ensured only in the climatic conditions specified in this manual and if all standard safety measures as well as the specific measures described in this manual are followed.

Do not use the instruments in places where there are:

- Corrosive or flammable gases.
- Direct vibrations or shocks to the instrument.
- High-intensity electromagnetic fields, static electricity.

### User obligations

The instrument operator shall follow the directives and regulations below that refer to the treatment of dangerous materials:

- EU directives on workplace safety.
- National law regulations on workplace safety.
- Accident prevention regulations.

## 11 Accessories ordering codes

The pyranometer is supplied with solar radiation protection screen, M12 female free connector (only if the optional cable is not ordered), calibration report and **DATAsense** PC application software downloadable from the website.

**Cables and fixing accessories must be ordered separately.**

### Fixing accessories

---

- |             |  |
|-------------|--|
| <b>LPS1</b> | Fixing bracket for Ø 30...50 mm mast. Installation on horizontal or vertical mast.           |
| <b>LPS5</b> | Adjustable holder for mounting the pyranometer in an inclined position on Ø 30...50 mm mast. |

### Installation cables

---

- |                   |   |
|-------------------|---|
| <b>CPM12-5...</b> | Cable with 5-pole M12 connector on one end, open wires on the other end. Length 5 m (CPM12-5.5) or 10 m (CPM12-5.10). |
|-------------------|---|

### PC connecting cable

---

- |                |  |
|----------------|--|
| <b>CP24K-5</b> | Kit for connecting the pyranometer RS485 output to a PC. It includes the SWD10 power supply and the RS485/USB adapter with: <ul style="list-style-type: none"><li>○ 5-pole M12 connector for connecting the pyranometer.</li><li>○ A-type USB connector for connecting the PC.</li><li>○ Jack connector for connecting the SWD10 power supply.</li></ul> |
|----------------|--|

### Spare parts

---

- |               |   |
|---------------|---|
| <b>LPSP1A</b> | UV-resistant solar radiation protection screen. |
| <b>LPG</b>    | Silica-gel (5 tablets).                         |

## NOTES

---

## **WARRANTY**

The manufacturer is required to respond to the "factory warranty" only in those cases provided by Legislative Decree 6 September 2005 - n. 206. Each instrument is sold after rigorous inspections; if any manufacturing defect is found, it is necessary to contact the distributor where the instrument was purchased from. During the warranty period (24 months from the date of invoice) any manufacturing defects found will be repaired free of charge. Misuse, wear, neglect, lack or inefficient maintenance as well as theft and damage during transport are excluded. Warranty does not apply if changes, tampering or unauthorized repairs are made on the product. Solutions, probes, electrodes and microphones are not guaranteed as the improper use, even for a few minutes, may cause irreparable damages.


The manufacturer repairs the products that show defects of construction in accordance with the terms and conditions of warranty included in the manual of the product. For any dispute, the competent court is the Court of Padua. The Italian law and the "Convention on Contracts for the International Sales of Goods" apply.

## **TECHNICAL INFORMATION**

The quality level of our instruments is the result of the continuous product development. This may lead to differences between the information reported in the manual and the instrument you have purchased.

We reserve the right to change technical specifications and dimensions to fit the product requirements without prior notice.

## **DISPOSAL INFORMATION**



Electrical and electronic equipment marked with specific symbol in compliance with 2012/19/EU Directive must be disposed of separately from household waste. European users can hand them over to the dealer or to the manufacturer when purchasing a new electrical and electronic equipment, or to a WEEE collection point designated by local authorities. Illegal disposal is punished by law.

Disposing of electrical and electronic equipment separately from normal waste helps to preserve natural resources and allows materials to be recycled in an environmentally friendly way without risks to human health.



**RoHS**

**senseca.com**



Senseca Italy S.r.l.  
Via Marconi, 5  
35030 Selvazzano Dentro (PD)  
ITALY  
[info@senseca.com](mailto:info@senseca.com)



**NIEUWKOOP**